SUPPLEMENTAL/BID BULLETIN NO. 2

Date: November 10, 2026
Title: IB-2026-10-28 SUPPLY, DELIVERY, AND INSTALLATION OF VARIOUS LABORATORY SUPPLIES AND EQUIPMENT
Reference No.: 111847

This bulletin is being issued to revise/clarify certain portions of the bidding documents. This shall form an integral part of the bidding documents for the above-mentioned procurement project.

QUERIES/REFERENCE	AMENDMENTS/CLARIFICATIONS
For Lots 3, 4, 6, and 7, the Technical Specifications require calibration and validation certificates. May we clarify whether the validation certificate should be issued by the manufacturer or the distributor? Further, may we submit either a calibration certificate or a validation certificate, or are both documents required?	Both the calibration and validation certificates are required and shall be issued by the manufacturer.
For Lot 7, Item 5 (Fume Hood), the specified external dimensions are 1,200 × 800 × 2,515 mm. May we clarify which measurements correspond to the width, depth, and height? Also, where will the equipment be installed, specifically on the ground floor or second floor? Further, does the specified dimension of 300 for the PVC exhaust duct refer to its diameter? May we also request clarification on the required length of the exhaust duct?	The correct external dimensions are 1,200 mm (W) × 800 mm (D) × 2,515 mm (H). The equipment will be delivered to the Carigara Campus. However, the specific building and installation location have not yet been determined. Installation will likely be on either the ground floor or second floor, subject to confirmation. The specified dimension of Ø300 refers to the diameter of the exhaust duct.
For Lot 6, may we request clarification regarding the specified pH accuracy requirement? Is the indicated accuracy of 10.1 correct, or is this a typographical error?	The required accuracy is ±10.1.
For Lot 7, Item 2 (Conductivity Meter), we would like to clarify the specified conductivity range of 0.0001 µS/cm to 1,000 mS/cm, as this appears to be an unusually wide range. May we be allowed to offer a unit with a conductivity range of 0–199.99 mS/cm? Further, regarding the temperature range, may we offer a unit with a temperature range of -10°C to 110°C and temperature accuracy ranging from ±0.1°C to ±0.4°C to accommodate our proposed brand?	The original Technical Specifications for Lot 7, Item 2 (Conductivity Meter), shall be retained. Bidders shall therefore comply with the prescribed conductivity range, temperature range, and temperature accuracy as indicated in the Bidding Documents.
For Lot 7, Item 6, the specification indicates a preference for 316 stainless steel. May we clarify whether other grades of stainless steel, such as 304 stainless steel, may also be offered?	The work surface shall be made of stainless steel with a grade ranging from 304 to 316.
For Lot 7, Item 7 (Analytical Balance), may we request that the settling time requirement be adjusted from 2 seconds to 3 seconds, subject to the end-user's approval? Further, considering that different models of analytical balances have varying minimum sample weight specifications, may the requirement of approximately 16 mg at 1% tolerance be reconsidered and instead be stated as "minimum sample weight as specified by the manufacturer for the offered model" to allow flexibility in the brands and models that may be offered?	The settling time shall be 2–3 seconds, while the minimum sample weight shall be as specified by the manufacturer for the offered model, supported by official technical documentation.
For Lot 7 (Analytical Balance), the calibration requirement states: "supplied with factory calibration certificate and/or provision for calibration upon delivery." May we clarify whether compliance with either of the two requirements is sufficient, or whether both are required? Will	Both requirements shall be complied with. The phrase "and/or" shall be amended to "and" to clearly indicate that both requirements are mandatory.

QUERIES/REFERENCE	AMENDMENTS/CLARIFICATIONS
compliance with only one of these requirements affect the evaluation of the bid?	
For Lot 9 (Thermal Imaging Camera), may we request clarification regarding the specified temperature accuracy? Is the required accuracy 2°C or 2% of the reading?	The requirement is expressed with a ± (plus/minus) sign. Please refer to the amended Technical Specifications in Annex "A".
With respect to the Single Largest Completed Contract (SLCC) requirement, will the required contract amount be based on the ABC of the entire project or only on the ABC of the lot or lots being participated in?	The SLCC requirement shall be based only on the ABC of the lot or lots being participated in.
May we request an extension of the required delivery period from sixty (60) calendar days to ninety (90) calendar days?	Yes, the delivery of the Goods shall be completed within ninety (90) calendar days from receipt of the Notice to Proceed (NTP), instead of the original sixty (60) calendar days.
May we request clarification regarding the specific campuses or delivery locations for the items under this project?	The detailed delivery locations and arrangements shall be discussed further with the identified Lowest Calculated Bidder (LCB) during the post-qualification process.

All statements and formats referring to this clause should be amended/corrected accordingly.

For guidance and information of all concerned.

For further information, please refer to:

(SGD) VINCENT B. CABANTOC
Head, BAC Secretariat
CP No. 0953-355-7046 - TM
Email Add: evsu.bacsecretariat@evsu.edu.ph

Noted:

(SGD) DANILO B. PULMA, DM
Vice President for Administration, Finance, & Auxiliary Services
Chairperson, Bids and Awards Committee

Section VII. Technical Specifications

IB-2026-10-28 SUPPLY, DELIVERY, AND INSTALLATION OF VARIOUS LABORATORY SUPPLIES AND EQUIPMENT				
Lot No.	**Unit**	**Item Description**	**Qty**	**Statement of Compliance**
1	**LOT**	**09-0601-26-IPDS**	**1**	
1	unit	**Supply, delivery, and testing of Compact Electric Fish Feed Pellet Machine**  *Capacity: 20-40 kg/hour (minimum) *Motor Power: 2.2-3.0 kW *Power Supply: 220V, Single Phase, 60 Hz *Pellet Diameter: 1 mm , 2 mm, 3 mm, 4 mm, interchangeable dies *Pellet Length: Adjustable *Feed Type: Suitable for sinking and semi-floating fish feed *Construction: Stainless 316 *Feeding System: Hopper with manual or automatic feeding *Pellet Moisture Requirement: 10-15% *Includes: - Four (4) pellet dies - Basic took kit - User's manual - Spare roller * Safety Features: - Emergency stop switch - Overload protection -Protection safety cover *Warranty: Minimum one (1) year	1	
2	**LOT**	**09-0602-26- IPDS**	**1**	
1	unit	**Supply, delivery, installation, and commissioning of Vertical Steam Pressure Autoclave**  *Capacity: 35 Liters (minimum) *Chamber Material: SUS316 Stainless Steel Sterilization *Temperature: 105˚C-134˚C *Working Pressure: 0.22 MPa *Temperature Accuracy: ±1˚C *Timer: 0-60 minutes or better *Digital temperature and timer display *Electric heating system *Stainless steel sterilization baskets (minimum two) *Pressure gauge *Safety valve *Over-temperature protection *Over-pressure protection *Low-water level protection *Automatic shut-off after sterilization *Power Supply: 220 , Single Phase, 60 Hz *Complete with accessories, operation manual, and power cord *Warranty: Minimum one (1) year	1	
3	**LOT**	**09-0603-26-IPDS**	**1**	
1	unit	**Supply, delivery, and testing of Portable handled chlorophyll meter**  *Non-destructive measurement of leaf chlorophyll content *Measuring Range: -9.9 to 199.9 SPAD Units *Display Resolution: 0.1 SPAD Unit *Measuring Accuracy: ±1.0 SPAD Unit (0.0-50.0 SPAD) *Repeatability: ±0.3 SPAD Unit *Measuring Area: 2 mm x 3 mm *Light Source: Two LED light sources *Detector: Silicon photodiode *Measurement Time: Less than 2 seconds *Memory Capacity: Minimum 30 measurements *Statistical Functions: -Average -Standard Deviation -Data Count	2	

		*LCD Digital Display *Automatic Power-Off *Battery: Two (2) AA alkaline batteries *Portable and lightweight design *Complete with carrying case *User manual *Battery set *Calibration and Validation Certificate *Warranty: Minimum one (1) year		
4	**LOT**	**09-0604-26-IPDS**	**1**	
1	unit	**Supply, delivery, and testing of Portable waterproof multiparameter water quality meter** *Measures: -Ph -Dissolved Oxygen (DO) -Electrical Conductivity (EC) -Total Dissolved Solids (TDS) -Salinity -Temperature *Automatic Temperature *Compensation *Large backlit LCD display *Data logging capability *USB data transfer *Automatic calibration *Replaceable battery *Complete calibration solutions *Complete sensor/probe assembly *Power adapter and USB cable *User Manual *With Calibration Certificate or Any Validation Certificates 	1	
5	**LOT**	**09-0605-26-IPDS**	**1**	
1	unit	**Supply, delivery, installation, and testing of Digital laboratory ultrasonic cleaner** Capacity: 10 liters minimum. Ultrasonic Frequency: 40 kHz. Ultrasonic Power: 400 W or higher. Heating Function: Yes. Temperature Range: Ambient to at least 80˚C. Digital Timer: 0-60 minutes or better. Digital control of Temperature and timer. Stainless-steel cleaning tank. Stainless-steel cover/lid. Stainless-steel wire basket included. Drain valve for convenient removal of cleaning solution. Suitable for cleaning laboratory glassware, instruments, metal components, electronic components, and other small laboratory parts. Electrical Requirement: 220-240 V AC, 50/60 Hz. Complete with power cord and standard accessories. User/operation manual included. Installation/testing upon delivery. With validation certificates. Minimum one (1) year warranty 	2	
6	**LOT**	**09-0606-26-IPDS**	**1**	
1	Lot	**Supply, delivery, and testing of Digital soil nutrient and fertilizer analyzer, laboratory/portable type** Application: Academic laboratory, soil analysis, agricultural research and instructional activities Parameters: Nitrogen (N), phosphorus (P), potassium (K), pH, soil moisture and salt/electrical conductivity  Additional Nutrient Analysis: Capability for selected macro- and micronutrients, preferably including Ca, Mg, S, Fe, Mn, B, Zn, Cu and other applicable elements Fertilizer Analysis: Capable of analyzing chemical, organic, foliar, water-soluble and spray fertilizers Organic Matter. Measurement capability preferred Heavy Metal Screening: Capability for selected beavy metals such as Pb, Cr, Cd, Hg and As, where applicable	1	

		Measurement Principle: Photoelectric/colorimetric or equivalent analytical method Measurement Range: Approximately 0.001-9999 or equivalent, depending on parameter pH Range: 1-14 pH Resolution: 0.01 pH pH Accuracy/Error. Approximately ±10.1 pH or better Salt/Conductivity: Approximately 0.01%-1.00% Salt/Conductivity Error: Approximately +5% relative error or better Soil Moisture: 0-100% Moisture Error: <0.5% Resolution: 0.001 or better Display. Digital LFD/LCD display Control: Microprocessor/microcomputer-controlled Data Output: Digital data display and/or printed test results Printer. Built-in thermal printer preferred Test Capacity: Capability to analyze individual and multiple soil samples Sample Preparation: Compatible with standard soil extraction/preparation procedures Power Supply: 110-240 V AC, 50/60 Hz or suitable for Philippine 220 V/60 Hz Power. AC/DC operation preferred Accessories: Required sample holders, extraction containers, cuvettes/test vessels, power adapter, printer paper and standard accessories Reagents: Initial set of required soil/fertilizer testing reagents sufficient for initial laboratory use Calibration: Calibration/verification capability and instructions included Documentation: User manual and technical documentation Calibration and Validation Certificate: Included, where applicable Warranty. Minimum one (1) year Installation/Testing: Delivery, installation/setup. Operational testing and user orientation included		
7	**LOT**	**09-0609-26-IPDS**	**1**	
1	lot	**Supply, delivery, installation setup, testing, and calibration of Inverted Fluorescence & Phase Contrast Tissue Culture Microscope.** Type: Laboratory-grade inverted microscope suitable for tissue/cell culture.  Optical System: Infinity-corrected optical system or equivalent Observation Modes: Brightfield, Phase Contrast, and Epi-Fluorescence Illumination: LED transmitted-light illumination Fluorescence Illumination: LED or equivalent solid- state fluorescence illumination Fluorescence: Mininaun 2 fluorescence filter sets/cubes, preferably DAPI and FITC or equivalent Objectives: Minimum 4, 10-, 20 and 40-. including phase-contrast objectives Phase Contrast: Available at mininmun 10, 20 and 40- Eyepieces: WF10 or equivalent, minima 22- field mumber Condenser: Long-working-distance condenser suitable for tissue/cell culture vessels Stage: Mechanical stage suitable for Petri dishes culture plates, slides and/or multi-well plates Specimen Holders: Suitable holders for Petri dishes. culture plates and microscope slides Focusing: Coaxial coarse and fine focusing mechanism Camera Port: Trinocular phototube or equivalent camera attachment capability Camera: Digital color camera, preferably 25 MP. with USB/computer connectivity Image Documentation: Capable of capturing still images and video Software: Basic image acquisition/documentation software included Interpupillary Distance: Adjustable Power: 100-240 V AC, 50/60 Hz or suitable for Philippine 220 V supply Accessories: Dust cover, necessary power cables, filters, adapters, specimen holders and standard accessories	1	

		Warranty: Minimum one (1) year Certification: With Calibration/ Validation Certificates Installation: Delivery, installation, testing. calibration and user orientation included		
2	set	**Supply, delivery, installation setup, and testing of Benchtop pH and Conductivity Meter. Two- Chamel, Laboratory Type** Type: Digital benchtop laboratory pH and conductivity meter. Measurement: Simultaneous or dual-channel pH and conductivity. pH Range: Approximately -2.000 to 19.999 PH pH Resolution 0.001/0.01/0.1 pH pH Accuracy: 0.002 pH or better my Range: Approximately -1.999 to -1.999 mV Conductivity Range: Approximately 0.001 µS/cm to 1,000 mS/cm Conductivity Accuracy: 0.5% or better Temperature Range: Approximately -30°C to 130°C Temperature Resolution: 0.1°C Temperature Accuracy: 0.1°C Temperature Compensation: Automatic/manual Temperature compensation TDS: Measurement calculation capability Salinity: Measurement/calculation capability Resistivity: Measurement/calculation capability Calibration: Multi-point calibration with automatic calibration recognition Data Management: GLP-compliant measurement/data documentation Display: Large digital LCD display showing measurement and temperature Inputs: Compatible with pH electrode, conductivity electrode and temperature sensor Connectivity: USB and/or RS-232 or equivalent data-transfer interface Software: Basic PC data-transfer/documentation capability, if applicable Power: 100-240 V AC, 50/60 Hz or suitable for Philippine 220 V supply Electrode Stand: Included Protective Cover: Included pH Electrode: Included Conductivity Electrode: Included Temperature Sensor: applicable probe Included or integrated with Calibration Solutions: Initial pH and conductivity calibration solutions included User Manual: Included Test Calibration Certificate: Included Warranty: Minimum one (1) year Installation Testing: Delivery, setup and operational testing included 	1	
3	lot	**Supply, delivery, installation/setup, calibration, and/or commissioning of Automatic Self-Cleaning Laboratory Water Distillation Unit.** Laboratory-grade automatic water distillation unit. Minimum distillation capacity: 100 L/24 hours. Distilled water quality: 1-2 µS/cm or better at 25°C. Adjustable heating power: 750-3,000 W or better. Power supply: 208-240 VAC, 50-60 Hz. Energy consumption: 0.7 kWh/L or better. Suitable for continuous/24-hour operation. Microprocessor-controlled operation. Touchscreen or digital control display. Automatic water-level control. Automatic distillation process control. Automatic self-cleaning system. Built-in cleaning-agent reservoir. Automatic sterilization function. Automatic protection against dry running and overheating. PID-controlled cooling-water management or equivalent. Integrated heat exchanger for water/energy conservation. Water-contact components made of 316 stainless steel or equivalent laboratory-grade corrosion- resistant material. 	1	

		Compact benchtop design. Distillate dispensing outlet. Appropriate inlet-water pressure control. Calibration/ validation, and safety certifications. Complete with necessary tubing, fittings, power cord, accessories, and cleaning-agent reservoir.		
4	lot	**Supply, delivery, installation/setup, calibration, testing, and/or commissioning of Digital single-beam UV-Visible spectrophotometer.**  Wavelength Range: 190-1,100 nm or wider. Spectral Bandwidth: 2 nm Wavelength Accuracy: 1.0 nm or better. Wavelength Repeatability: 0.5 m Photometric Range. At least -0.3 to 3.0 Abs or equivalent. Photometric Accuracy: ±0.5% T or better. Photometric Repeatability: 50.2% T Optical System: Single-beam grating monochromator. Detector: Silicon photodiode or equivalent. Light Sources: Deuterium lamp for UV and tungsten- halogen lamp for visible range, or equivalent. Lamp Switching: Automatic or wavelength- dependent. Wavelength Selection: Digital/microprocessor- controlled. Measurement Modes: Absorbance, % Transmittance and concentration. Scanning: Wavelength scanning capability preferred. Sample Holder: Standard 10-mum cuvette holder. Cuvettes: Minimum set of compatible quartz and optical-glass/plastic cuvettes Display: Digital LCD or color display Data Storage: Internal memory or USB data storage. Connectivity: USB and/or RS-232 Software: Basic PC data acquisition/documentation software, if applicable. Calibration: Automatic wavelength/photometric calibration or equivalent. Power Supply: 100-240 V AC, 50/60 Hz. Accessories: Dust cover, power cord, cuvette holder. cuvettes and standard accessories User Manual: Included Calibration and Validation Certificates. Warranty: Minimum one (1) year Installation/Testing: Delivery, installation, operational testing and user orientation included	1	
5	lot	**Supply, delivery, assembly/ installation, testing, and calibration of Ducted Chemical Fume Hood, Laboratory.**  approximately 1,200 mm width: single-user work station; external dimensions approximately 1,200mm (W) x 800mm (D) x 2,515mm (H) or equivalent; internal working area approximately 990mm (W) × 600 mm (D) × 870 mm (H) or equivalent; work surface height approximately 900 mm: maximum sash opening approximately 750 mm; adjustable face/air velocity of 0.3-0.8 m/s; exhaust volume approximately 700 m/h or higher; built-in corrosion-resistant centrifugal blower; adjustable blower speed; chemical-resistant phenolic resin worktop: 5-mm toughened/tempered glass vertical sash, height- adjustable: LED work lighting: PVC exhaust duct, approximately Ø300 mm, with necessary ducting and pipe straps: chemical-resistant construction: water sink and water tap: electrical outlets; base cabinet: microprocessor/digital control system; visual/audible alarm for unsafe operating conditions where applicable: 220 VAC, 50/60 Hz. complete with installation, testing and commissioning: user manual; minimum one (1) year warranty: with certifications such as validation/ safety certificate	1	

6	lot	**Supply, delivery, installation, and commissioning/testing of Vertical Laminar Flow Cabinet-Laboratory Type.** Type: Vertical laminar airflow cabinet/clean bench. Configuration: Vertical, unidirectional downward airflow,  User Capacity: Single user Working Width: Approximately 1,200 mm/4 ft. External Dimensions: Approximately 1,300 × 700-750x2,040 mm or equivalent. Internal Working Area: Approximately 1,200 × 600-650x600-650 mm. Working Surface Height: Approximately 750 mm. Airflow Velocity: 0.3-0.5 m/s, adjustable. Airflow: Vertical, unidirectional. Cleanliness: ISO Class 5/Class 100. HEPA Filter: Minimum 99.995% efficiency at 0.3 um. Pre-filter: Washable polyester-fiber or equivalent Work Surface: from 304 stainless steel to 316 stainless steel. corrosion-resistant Main Body: Cold-rolled steel with anti-bacterial powder coating or equivalent Front/Side Windows: Minimum 5-mm toughened/tempered glass, preferably anti-UV Front Window. Height-adjustable motorized or equivalent Maximum Opening: Approximately 350-430 mm Lighting: LED, minimum 300 lux UV Lamp: Germicidal UV-C lamp, approximately 253.7 nm, with timer/interlock Display: Digital/LCD display with microprocessor control Airflow Control: Adjustable airflow speed Noise Level: ≤65 dB(A) Vibration 55 um RMS, where applicable Electrical Outlet: Minimum two waterproof electrical sockets Power Supply: AC 220 V 10%, 50/60 Hz Power Consumption: Approximately 400-700 W. depending on configuration Base Stand: Included Caster: Universal caster with leveling feet Safety: Airflow monitoring/alarm and UV/blower interlock Accessories: LED lamp, UV lamp, base stand, electrical outlets and other standard accessories Installation: Delivery, installation, testing and commissioning included Documentation: User/operation manual and technical documentation Certification: With validation/ safety certifications Warranty: Minimum one (1) year	1	
7	lot	**Supply, delivery, setup, testing, and calibration of Analytical Balance.**  Maximum Capacity:220 g Readability: 0.1 mg (0.0001 g) Repeatability: 0.1 mg Linearity Deviation: 0.2 mg Settling Time: 2–3 seconds Calibration/Adjustment: Internal adjustment/calibration Weighing Pan: approximately 90 mm diameter Draft Shield: Included Display: Digital display Interface: RS-232 or equivalent digital communication interface Minimum Sample Weight: As specified by the manufacturer for the offered model, supported by official technical documentation Power Supply: 220-240 VAC, 50/60 Hz, or compatible with Philippine electrical standards Construction: Laboratory-grade, durable housing suitable for routine academic laboratory use Accessories: Draft shield, weighing pan, power adapter/cord, operating manual, and calibration documentation Warranty: At least one (1) year Calibration: Supplied with factory calibration certificate and provision for calibration upon delivery	1	

8	LOT	09-0610-26 IPDS	1	
1	unit	**Supply, delivery, and testing of Handheld Multi- Band Multi-GNSS GPS Receiver.**  Type: Rugged handheld GPS/GNSS receiver GNSS Support: Multi-GNSS and multi-band satellite reception Satellite Systems: GPS, GLONASS, Galileo, QZSS and equivalent supported GNSS Multi-Band: Required for improved positioning accuracy in challenging environments Receiver: High-sensitivity GNSS receiver Display: Minimum 2.6-inch color display Display Resolution: Minimum 160 x 240 pixels or better Display: Sunlight-readable/transflective color display Electronic Compass: 3-axis, tilt-compensated electronic compass Altimeter: Barometric altimeter Positioning: Waypoint, track and route recording Waypoints: Minimum 5,000 Tracks: Minimum 200 saved tracks Track Log: Minimum 10,000 points Routes: Minimum 200 Arca Calculation: Required Coordinate Systems: Support for standard geographic coordinate systems and datums Coordinate Formats: Latitude/Longitude, UTM/UPS and other standard formats Map Compatibility: Ability to load/add maps Preloaded Maps: Topographic or equivalent base maps preferred Custom Maps: Capability to use custom maps preferred Data Transfer: USB interface Wireless Connectivity: Bluetooth and/or ANT+ preferred Mobile Application: Compatible with a GPS/GNSS field-data management application Water Resistance: Minimum IPX7 Battery: Replaceable/rechargeable battery system Battery Life: Minimum 16 hours under normal GPS operation Memory: Minimum 8 GB, preferably 16 GB or higher Field Functions: Waypoint averaging, track recording, route navigation, area calculation and distance measurement Georeferencing: Capability to record geographic coordinates and export location data Data Export: GPX or equivalent standard geographic data format Accessories: USB cable, battery/batteries, carabiner/hand strap, carrying case and documentation Warranty: Minimum one (1) year Training: Basic user orientation included	4	
9	**LOT**	**09-0611-26 IPDS**	**1**	
1	unit	**Supply, delivery, and testing of Handheld Thermal Imaging Camera.**  Type: Professional handheld thermal imaging camera Application: Building/facility inspection, infrastructure assessment, electrical/mechanical inspection, energy-efficiency assessment and field survey Infrared Resolution: Minimum 240 x 180 pixels Thermal Sensitivity (NETD): <50 mK Temperature Measurement Range: Approximately -20°C to 550°C Temperature Accuracy: ±2°C or ±2% of reading, or better Detector: Uncooled microbolometer Spectral Range: Approximately 7.5-14 jum Field of View. Approximately 33° x 25° or equivalent Minimum Focus Distance: Approximately 0.5 m or better Image Frequency: Minimum 9 Hz Digital Camera: Minimum 5 MP Image Enhancement: Thermal/digital image overlay or equivalent Image Modes: Thermal, digital, picture-in-picture and/or equivalent Color Palettes: Multiple selectable thermal color palettes	1	

		Display: Minimum 3.5-inch color touchscreen Display Resolution: Minimum 640 x 480 pixels Measurement Functions: Conter spot, area/box, minimum/maxximum temperature and temperature difference Emissivity Correction: Adjustable Reflected Temperature Correction: Required Atmospheric Compensation: Preferred Alarm/Isotherm: Above/below temperature alarm preferred Image Storage: Internal memory Connectivity. Wi-Fi and/or USB Data Transfer. Ability to export thermal images and measurement data Reporting: Compatible with manufacturer's thermal-analysis/reporting software or equivalent Battery: Rechargeable lithium-ion Battery Operation: Minimum approximately 4 hours continuous operation Environmental Protection: Minimum IP54, preferably IP54 or better Operating Temperature: Suitable for normal field inspection conditions Drop Resistance: Minimum approximately 2 m preferred Accessories: Rechargeable battery, charger/power supply, USB cable, carrying case and standard accessories Software: Basic thermal image analysis/reporting software, if applicable Calibration: Factory calibration certificate/documentation Warranty. Minimum one (1) year Training: Basic user orientation/operation demonstration included		